

Certificate of Analysis(COA)



HQ
12F, 5, Digital-ro 26-gil, Guro-gu, Seoul,
08389, Korea
Tel: +82-2-6220-7728
Fax: +82-2-6220-7721

Manufacturing Site
851-14, Seohae-ro, Paltan-myeon,
Hwaseong-si, Gyeonggi-do,
18531, Korea
Tel: +82-31-350-0807
Fax: +82-31-350-0808

US Operation
NanoEnTek America, Inc.
220 Bear Hill Road, Suite 102,
Waltham, MA 02451, USA
Tel: +1-781-472-2558
Fax: +1-781-790-5649

Website.
www.nanoentek.com

e-mail.
sales@nanoentek.com

Product Name

ADAM-rWBC Kit

Model Number

AD1K-050

Lot Number

3RW16861

Date of Manufacture

December 16, 2020

Expired Date

December 15, 2021

Components

Product Name	Model Number	Lot Number
r-Slide	RDS-50	3W16881, 3W16812
r-Solution	RDR-50	1RS16960
Standard Bead Solution	ADST-001	2SB16861 – 151 ~ 195

Certificate of Analysis(COA)

QC List

r-Slide, Chamber Height

Specification: $200 \pm 10 \mu\text{m}$

Result: Pass

r-Solution

pH $5.6 \pm 7\%$

Absorbance: Pass

Standard Bead Solution

Specification: Check the concentration on each bottle label.

Result: Pass

Certified Date : January 27, 2021

Certified by : 

Jason Jeong,

Quality Management Representative